



DECLARATION OF EC CONFORMITY

Manufacturer Akces-MED Ltd.

Declare that standing frame GENIUS

BASIC UDI-DI: 59038165GNS27
MODELE: GENIUS 1, GENIUS 2, GENIUS 3

marked with the CE sign is Class I medical device, of rule 1, consistent with the requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.

The above product to which this declaration applies complies with Regulation (EU) 2017/745 of the European Parliament and of the Council.

Conformity assessment was carried out in accordance with the requirements of the Regulation (EU) 2017/745 of the European Parliament and of the Council.

In order to demonstrate the safety and effectiveness of the medical device, the following standards were used to assess conformity:

PN-EN 1041+A1:2013-12

Information provided by the producer together with the medical product.

PN-EN 12182:2012

Assistive products for persons with disability - General requirements and research methods.

PN-EN ISO 15223-1:2017-02

Medical devices Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements.


PN-EN ISO 13485:2016

Medical products - Quality management systems - Requirements for law regulations aims.

PN-EN ISO 14971:2020

Medical devices - Application of risk management to medical devices.

President of the board



Sławomir Wroński