



SPECIAL NEEDS STROLLER
T A T A L U™



USER MANUAL





Manufacturer takes responsibility for proper usage of the device only when the product was purchased from qualified **Akces-MED Ltd.** employee or at specialized medical store!



Akces-MED Ltd. reserve the right to introduce technical and trading changes in the content of the instruction without warning.



Product marked with the CE sign is Class I medical device, of rule 1, consistent with the requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.



Akces-MED Ltd. introduced quality management system ISO 13485 regarding production, sales and service. The system is certified by Det Norske Veritas.





CONTENTS

| | |
|--|----|
| 1. The purpose of user manual..... | 5 |
| 2. Indications and contraindications for use..... | 5 |
| 3. General characteristics of stroller..... | 6 |
| 3.1 Standard equipment..... | 6 |
| 3.2 Accessories..... | 7 |
| 4. Size table..... | 9 |
| 5. How to use the stroller..... | 10 |
| 5.1 Unfolding the stroller..... | 11 |
| 5.2 Folding the stroller..... | 12 |
| 5.3 Adjustable footrest..... | 13 |
| 5.4 Seat depth adjustment..... | 14 |
| 5.5 5-point safety belts assembly..... | 14 |
| 5.6 Push handles adjustment..... | 17 |
| 5.7 Carrying handle for a folded stroller assembly..... | 17 |
| 6. Accessories..... | 19 |
| 6.1 TTL_001 LED lights..... | 19 |
| 6.2 TTL_013 Safety rail..... | 19 |
| 6.3 TTL_101 Thigh abduction belts..... | 19 |
| 6.4 TTL_104 Adjustable headrest..... | 20 |
| 6.5 TTL_114 H harness/ TTL_130 4-point safety vest..... | 21 |
| 6.6 TTL_128 Abduction block..... | 22 |
| 6.7 TTL_161 Foot and ankle stabilizing sandals..... | 22 |
| 6.8 TTL_162 Dynamic ankle stabilizers (FP-22)..... | 22 |
| 6.9 TTL_402 Umbrella..... | 23 |
| 6.10 TTL_403 Tray..... | 23 |
| 6.11 TTL_405 Folding canopy..... | 24 |
| 6.12 TTL_408 Rain cover / TTL_409 Mosquito net..... | 25 |
| 6.13 TTL_411 Cup holder..... | 25 |
| 6.14 TTL_417 Winter footmuff with muffs / TTL_447 Winter muffs / TTL_445 Footmuff..... | 26 |
| 6.15 TTL_448 Self-leveling tray..... | 26 |
| 6.16 TTL_501 Bag / TTL_509 Organizer..... | 27 |
| 6.17 TTL_505 Under seat storage basket..... | 27 |
| 6.18 TTL_506 Carrier bag for the stroller..... | 27 |
| 7. The use of stroller..... | 28 |
| 7.1 Preparation of the user..... | 28 |
| 7.2 Preparation of the stroller..... | 29 |
| 7.3 How to use vests..... | 30 |
| 8. Safety rules..... | 33 |
| 9. Term of warranty..... | 35 |
| 10. Maintenance and cleaning..... | 35 |
| 11. Labelling..... | 39 |
| 11.1 Symbols..... | 39 |
| 11.2 Sample data plate..... | 39 |





INTRODUCTION

TATALU™ – a stroller which makes walks with your child even more comfortable. A new version of the stroller with a lightweight and durable construction designed to perform well in all conditions. Its umbrella frame folds easily and quickly and takes up little space after folding. A number of improvements enhance the comfort of use both for the child and caregiver. The stroller can be fully customized – you decide which accessories will work for you.

TATALU™ – means comfort and stabilization! A number of improvements, such as the SOFT upholstery covering the sides of the stroller to provide protection of the user, adjustable 5-point safety belts, headrest or footrest with the possibility of changing the height, make the young user feel comfortable and safe. The SOFT upholstery meets the requirements of caregivers, which is confirmed by the CIGARETTE TEST proving that it is made of a non-flammable fabric. The increased comfort is due to an independent, reinforced shock absorption system and swivel front wheels which provide great maneuverability of the stroller.

TATALU™ – see for yourself that walks with your child can be even more enjoyable!

Remember to read this user manual carefully. You will find not only useful information but also guidelines safe and effective use of the product.

If you have any questions or remarks, do not hesitate to contact with us:

Medical adviser: mobile +48 516 183 883

Akces-MED Ltd.: ph. +48 17 864 04 70

e-mail: export@akces-med.com



Before using the product, user is obliged to familiarize himself/herself with user manual. Remember that complying with guidelines is highly important. It assures safety and extends durability as well as aesthetics of the device.





1. THE PURPOSE OF USER MANUAL



This user manual contains basic informations necessary to properly: prepare the stroller before use, use the stroller, maintain and clean the stroller.

This user manual should be kept in an easily accessible place. The manual is intended for people who look after disabled children, as well as for doctors and physiotherapists handling the stroller.

Before using the **TATALU™** stroller for the first time, the user is obligated to familiarize himself/herself with the user manual.

2. INDICATIONS AND CONTRAINDICATIONS FOR USE

These medical devices are intended for people with orthopedic, neurological and cardiological diseases and for people after surgery, with musculoskeletal dysfunctions, who have not acquired the ability to sit independently or who have difficulties in adopting and maintaining the sitting position.

Both for flaccid people with reduced muscle tone and for people with spasticity. The device can also be really useful for people who require long-term relief of one or both lower limbs. It is an irreplaceable aid in the rehabilitation process, both in terms of movement and social integration. The device is great help during everyday activities such as rehabilitation but also during playing, studying or having meals. The selection of the appropriate size of the device depends on medical indications and anatomical parameters, regardless of the age of the user.

This stroller is recommended for children in the following cases:

- Cerebral palsy – various types and forms,
- Brain and spinal cord injuries with lower limbs paralysis and weakness, balance and motor coordination disorders with different etiology,
- Demyelinating disease (e.g. multiple sclerosis - SM),
- Meningocele,
- Muscles diseases with weakness and paralysis (dystrophy, myasthenia gravis, myositis),
- Genetic disorders,
- Degeneration and neurological diseases,
- Other condition connected with muscles weakness, paralysis and musculoskeletal disorders,
- After head and spinal cord injuries and spine trauma with paralysis and paresis.

Contraindications for sitting:

- Critical condition of the user,
- Paralysis,
- Deep sedation,
- Unstable fractures with specific orthopedic recommendations, including spine fractures,
- Untreated deep vein thrombosis,
- Hemodynamic instability,
- Identified brain dysfunction when sitting upright,
- Severe arthritis of the cervical spine,
- Pulmonary embolism,
- Recent myocardial infarction in the last 24 hours,
- Fresh postoperative wounds or non-healing wounds in the back area, pelvic girdle area, hip joints and thighs,
- Users with mental disorders should not be left in the device without the supervision of a caregiver.





3. THE PURPOSE OF USER MANUAL

3.1 STANDARD EQUIPMENT

- ① Foldable umbrella frame with seat depth adjustment
- ② Adjustable push handles
- ③ 5-point safety belts
- ④ High lateral supports
- ⑤ Adjustable footrest
- ⑥ SOFT upholstery covering the sides of the stroller
- ⑦ Front swivel PU wheels with direction blockade
- ⑧ Rear PU wheels with independent brakes
- ⑨ Carrying handle for a folded stroller



3.2 ACCESSORIES

| | | |
|---|---|--|
|  <p>TTL_001</p> <p>LED lights</p> |  <p>TTL_013</p> <p>Safety rail</p> |  <p>TTL_101</p> <p>Thigh abduction belts</p> |
|  <p>TTL_104</p> <p>Adjustable headrest</p> |  <p>TTL_114</p> <p>H harness</p> |  <p>TTL_128</p> <p>Abduction block</p> |
|  <p>TTL_130</p> <p>4 point safety vest</p> |  <p>TTL_161</p> <p>Foot and ankle stabilizing sandals</p> |  <p>TTL_162</p> <p>Dynamic ankle stabilizers (FP-22)</p> |
|  <p>TTL_402</p> <p>Umbrella</p> |  <p>TTL_403</p> <p>Tray</p> |  <p>TTL_405</p> <p>Folding canopy</p> |
|  <p>TTL_408</p> <p>Rain cover</p> |  <p>TTL_409</p> <p>Mosquito net</p> |  <p>TTL_411 TTL_452</p> <p>Cup holder / Phone holder</p> |



TTL_417
TTL_445

**Winter footmuff with muffs/
Footmuff**



TTL_447

Winter muffs



TTL_448

Self-leveling tray



TTL_501

Bag



TTL_505

Under seat storage basket



TTL_506

Carrier bag for folded stroller




TTL_509

Organizer





4. SIZE TABLE



SIZE TABLE

| Symbol | Measurement | Unit | Size 1 |
|----------------|--|------|----------------|
| W | indicative user height* | [cm] | 90 - 130 |
| A | seat width ** | [cm] | 38 (30) |
| L | seat depth | [cm] | 22, 30 |
| C ₁ | backrest height | [cm] | 60 |
| C | headrest height | [cm] | 45 - 60 |
| E | backrest width** | [cm] | 38 (30) |
| F | footrest length | [cm] | 12, 20, 28, 39 |
| | push handle height | [cm] | 104 |
| X | backrest tilt angle (relative to the seat) | [°] | 95 |
| Y | seat tilt angle | [°] | 25 |
| | maximum user weight | [kg] | 50 |

*The value cannot be main factor for selection of product size.

**High rigid trunk supports narrow the seat / backrest by 8 cm to 30 cm.

WEIGHT AND DIMENSIONS

| Measurement | | Size 1 |
|----------------------|------|--------|
| Width | [cm] | 62 |
| Width after folding | [cm] | 36 |
| Length | [cm] | 104 |
| Length after folding | [cm] | 120 |
| Height | [cm] | 104 |
| Height after folding | [cm] | 28 |
| Weight | [kg] | 13 |

NOTE: Measurement tolerance range is +/- 1 cm / ½ in.

NOTE: Approximate values, depending on the product configuration.



5. HOW TO USE THE STROLLER

KEY



Info icon - appears in the manual when you need to read the additional information.



Warning icon - appears in the manual when you need to take **special precautions**.



Incorrect product adjustment can lead to such side effects as fainting, body deformities, joint overload, etc. If the abovementioned symptoms occur, stop using the product immediately, take the user out of the product and lay them down. If the undesirable symptoms persist, contact a physician immediately. If the symptoms have disappeared, inform the attending physician about this fact in order to make a decision about the further rehabilitation process.







5.1 UNFOLDING THE STROLLER



Keep the user away from the moving parts when unfolding the stroller. There is a serious risk of injury.






Before placing the child in the stroller, make sure that the front strut and back strut are completely unfolded.

5.2 FOLDING THE STROLLER



Keep the user away from the moving parts when folding the stroller. There is a serious risk of injury.





5.3 ADJUSTABLE FOOTREST




To change the height of the footrest, press the levers under the footrest, firmly pull the footrest up and then attach it at the desired height. The footrest is height - adjustable.






5.4 SEAT DEPTH ADJUSTMENT





5.5 5-POINT SAFETY BELTS ASSEMBLY

i



Mount the belts at the height of your choice on the backrest of the stroller by selecting the holes in the upholstery at the same height.







The 5-point safety belts can be mounted in one of the 3 places on the seat of the stroller.



Always make sure that the user has the 5-point belts properly fastened while using the stroller.

In order to check that, pull the belts firmly in different directions and see if the belts do not unfasten.





5.6 PUSH HANDLES ADJUSTMENT



5.7 CARRYING HANDLE FOR A FOLDED STROLLER ASSEMBLY





6. ACCESSORIES

6.1 TTL_001 LED LIGHTS



48




Attach the LED lights to the stroller frame in the selected place.

6.2 TTL_013 SAFETY RAIL



49



50

6.3 TTL_101 THIGH ABDUCTION BELTS



51



52



Remove one of the sliding buckles located on the straps and follow the mounting instructions for the lower part of the 5-point safety belts (see page 15, pt. 5.5, fig. 22 - 29).






i

Replace the previously removed sliding buckle and put the strap through the strap buckle (fig. 53). Thread the end of the strap through the sliding buckle.

6.4 TTL_104 ADJUSTABLE HEADREST

i

The headrest can be mounted at 3 different heights.



6.5 TTL_114 H HARNESS TTL_130 4-POINT SAFETY VEST




To mount the upper belts of the vests, follow the mounting instructions for the upper part of the 5-point safety belts (see page 14, pt. 5.5 5-point safety belts assembly, fig. 20 - 21).




59



60



61



62



Replace the previously removed sliding buckle and the strap buckle. Thread the end of the strap through the strap buckle and then through the sliding buckle (fig. 62).



**Before using the vest, it is necessary to read the safety rules in the manual.
For information on how to use the vest safely, see pages 31 - 33
of the manual.**

**Improper use and failure to follow safety rules may endanger the health and
life of the user.**



6.6 TTL_128 ABDUCTION BLOCK



The adjustable abduction block can be mounted in one of the 3 places on the seat of the stroller.



63



64

6.7 TTL_161 FOOT AND ANKLE STABILIZING SANDALS



65



Thread the straps of the sandals through the holes under the bottom of the footrest, fasten the Velcro straps and check for proper fastening.

6.8 TTL_162 DYNAMIC ANKLE STABILIZERS (FP-22)




66



67







6.9 TTL_402 UMBRELLA



Clamp the umbrella to the frame of the stroller by tightening the knob firmly (**fig. 70**). The angle of the umbrella can be changed by pressing the button (**fig. 70**).







6.10 TTL_403 TRAY





6.11 TTL_405 FOLDING CANOPY






6.12 TTL_408 RAIN COVER TTL_409 MOSQUITO NET



To use a rain cover or mosquito net, it is necessary to unfold the canopy and attach the footrest.



6.13 TTL_411 CUP HOLDER






Attach the holder to the pin located on the frame of the stroller.





6.14 TTL_417 WINTER FOOTMUFF WITH MUFFS TTL_445 FOOTMUFF TTL_447 WINTER MUFFS



6.15 TTL_448 SELF-LEVELING TRAY



There is a bag filled with granules attached to the bottom of the tray with velcro. It has a special pocket which allows you to adjust the amount of the granules depending on your needs.

Attention! An additional bag with granules is available as an option.



6.16 TTL_501 BAG / TTL_509 ORGANIZER



91


i

Place the bag or the organizer on the stroller's push handles and secure them to the prevent slipping by using the fasteners or press studs on both sides of the stroller.


6.17 TTL_505 UNDER SEAT STORAGE BASKET




92



93



94



95

6.18 TTL_506 CARRIER BAG FOR THE STROLLER

i

Fold the stroller according to the folding instructions (see page 12, pt. 5.2, fig. 8 - 13) and place it in the cover. Make sure it is properly closed.



7. THE USE OF STROLLER



TATALU™ stroller has to be used with compliance to guidelines included in this user manual. Remember to use the product only with someone who is trained and knowledgeable.

Do not forget to prepare both product as well as user before using stroller!

7.1 PREPARATION OF THE USER

Before using the product, it is essential to make sure that:

- user's outfit is light and comfortable sporty clothes are advisable,
- shoes should be comfortable and, moreover, they should stabilize ankle,
- due to hygienic reason and to avoid abrasion it is advisable for the user to always wear clothes when using the device, skin must not be naked especially in places of regular contact with upholstery,
- drains, bags, catheters etc. if applied, have to be secured to function properly and not to disturb during using the product,
- if it is necessary for a user to wear orthopedic supply (collar, corset, orthosis, shoes) if it is recommended by doctor.



Outfit should protect user's skin from direct contact with upholstery as well as other parts of equipment.

Only doctor can make a decision concerning using any kind of orthopedic supply!





7.2 PREPARATION OF THE STROLLER

1. Before use it is necessary to:
 - unfold the stroller;
 - place the device on even ground;
 - make sure that all parts are serviceable, correctly mounted and stable;
 - block brakes;
 - adjust the device to user's measurements;
 - undo safety belts or vest so as to place easily the user in the device;
 - place user in the device;
 - fasten all elements which stabilize user e.g. vest, safety belts;
 - mount additional equipment e.g. tray.
2. In case of users who have problem with bladder, it is necessary to put special cloth on seat. This cloth can be purchased in pharmacy or medical shops.



In case of any obstacles such as step or curb, it is necessary to raise front wheels by pressing the handle down. Next, place front wheels on step and then raise back part of stroller with the aid of handle. Be careful not to tilt the stroller back too much. Remember to place stroller forwards.

It is essential to hold stroller with the aid of handle during using it otherwise stroller can roll.

Before the product is used, it is highly necessary to make sure that all belts as well as safety equipment are correctly fastened as well as adjusted. What is more, check if they do not become undone due to the pressure and, moreover, make sure that clampings from adjustable elements are firmly tightened!

Equipment works correctly only when the device is adjusted according to user's height as well as weight. Remember to take all these measurements into account while purchasing the product. Correct adjustment is the most important due to the fact that only then the product is safe. It is forbidden to force child to keep specific position because he/she will feel uncomfortable and it may lead to wrong posture.



7.3 HOW TO USE VESTS



The vest provides posture support for wheelchair users as well as users sitting in positioning chairs or other rehabilitation seats. When properly used, the vest allows the user to adopt the correct sitting position for a longer period of time. The design and shape of the vest make it possible for the user to be free to move their upper limbs within the constraints of the user.

If the vest is not used properly, it can pose a serious threat to the health and life of the user!

Possible threats incident to the use of the vest:

- 1) **If the lower vest straps are fastened too loosely, there is a risk that the vest will move upwards to the user's neck, making it difficult to breathe** – always check the tension of the straps and set them in such a way that the lower edge of the vest is located near the lower abdomen of the user.
- 2) **In a situation when the user slips off the seat, the upper edge of the vest cover's the user's neck and blocks the trachea, making it difficult to breathe** – it is therefore necessary to ALWAYS use the accessories stabilizing the user's pelvis, such as a pelvic belt, thigh abduction belts or an abduction block.
- 3) **There is a risk that the upper straps of the vest will press the user's neck, blocking the trachea and making it difficult to breathe. This might happen if the user's body is improperly secured in the seat or stroller** – it is then necessary to use the optional accessories of these devices and stabilize the user's body using the lateral supports. If there is no possibility to use such accessories with the device and the user's body cannot be properly stabilized, the device should be replaced with the one having such an option.
- 4) **The vest is fastened too tightly to the backrest and as a result the user cannot breathe easily** – in this case the vest should be loosened in such a way that it is possible for the caregiver to insert a straightened hand between the vest and the user's body. If, however, we want to stabilize the user's body properly, we should use a dynamic orthopedic vest, which does not cause the breathing difficulty for the user.



Instructions for the proper use of the vest:

- 1) Make sure that the user is sitting in the best possible position:
 - in the middle of the seat,
 - deeply in the seat in such a way that the user's lumbar spine rests against the backrest.
- 2) Secure the user's pelvis with, for example, a pelvic belt, thigh abduction belts or an abduction block.
- 3) Make sure that the side supports of the device, if used, are set in the correct position. It is required that the users with the poor trunk control use the side supports. The side supports limit the lateral movements of the user's body, which could lead to the compression of the vest straps on the neck of the user.
- 4) Secure the vest:
 - fasten and adjust the lower straps of the vest, so that the lower edge of the vest is located near the lower abdomen of the user,
 - fasten the horizontal strap of the vest (if present),
 - fasten the shoulder straps of the vest,
 - check if the buckles of the vest do not open under pressure.
- 5) Check the user's position:
 - make sure that the vest cannot move by itself towards the user's neck
 - make sure that you can insert a straightened hand between the vest and the user's body

To prevent injuries and to minimize the risks associated with the use of the vest, the following rules should be followed:

- 1) The position of the vest **MUST** be checked each time after putting the user in the seat.
- 2) Special attention should be paid to protecting the user from falling forward, slipping off the seat or lifting the pelvis. The vest **MUST** be used with the accessories which will secure the user's pelvis at all times (e.g. the pelvic belt, thigh abduction belts or the abduction block). The use of the vest cannot be considered separately but as a part of a general posture support system including supports such as pelvic belts, foot support, torso stabilization, thigh abduction belts, tilt-in-space position and reclining backrest.
- 3) Special attention should be paid to ensuring that the entire foot of the user has stable support. If the device is not equipped with a foot support, the user's feet must rest firmly against the ground.





- 4) It should be remembered that any change of the inclination angle of the backrest, seat, foot support or tilt-in-space function might affect the effectiveness of the belts and supports!
- 5) Special attention should be paid to the use of stabilization elements or side supports. If the user has poor trunk control, the side supports will limit the lateral movements, which could lead to the compression of the vest straps on the neck of the user.
- 6) The mounting of the vest and adjusting it to the user can only be carried out by a suitably trained person, who is familiar with the instructions for the use of the vest and the rules of its safe use.
- 7) Do not make any repairs or modifications to the vest on your own. In order to repair or replace any part of the vest, please contact the manufacturer.
- 8) Do not use a damaged or incomplete vest, as this poses a serious threat to the health and life of the user!
- 9) Do not leave the user in the device unattended!
- 10) The correctness of the use of the vest should be particularly controlled in accordance with the recommendations in case of children, patients with impaired memory or mentally ill.
- 11) Make sure that all people taking care of the user are familiar with the rules of the safe use of the vest.

How to care for the vest:

The vest works effectively only if kept clean and maintained in accordance with the manufacturer's instructions. Before each use of the vest, check whether:

- 1) The vest straps are not frayed or worn. If so, have them exchanged for new ones.
- 2) The vest straps are not attached to the wheels of the device.
- 3) The buckles are not damaged. If so, have them exchanged for new ones.
- 4) The velcro fastening holds firmly. If not, remove any dirt or hair from the velcro, or have it replaced with a new one.
- 5) There have been no changes in the user's height or weight from the moment the vest was adjusted to the user.
- 6) The shoulder straps of the vest are adjusted so that they are not too loose or too tight.
- 7) The vest should be cleaned regularly with the use of warm soapy water and dried out of the direct sunlight and other sources of heat. The vest must always be dry before usage.

If you have any doubts or questions about the use of the vest, please contact your doctor or physiotherapist.



8. SAFETY RULES



TATALU™ stroller has to be used in compliance with guidelines in this user manual.

It is highly important to consult either doctor or physiotherapist in order to make right decision concerning the choice of stroller and its size. It is also essential to familiarize with the guidelines described in user manual before using the device.

1. Before use it is necessary to:

- make sure that there are no damages or different faults which could endanger patient's life; check if all parts are working and are correctly mounted and, moreover, check if they are not broken,
 - check if all screws, knobs and all glued, sewed or tapped parts are tightened,
 - check if brakes are operating correctly,
 - check if buckles are operating correctly,
 - always use safety belts, velcro tapes should overlap min. 8cm,
 - adjust the device according to patient's measurements and needs e.g. the adjustment of pelottes, footplate etc.,
 - make sure that all movable parts and knobs for adjustments are stable and locked,
 - secure patient with the aid of elements such as vest, foot stabilizing belts, pelottes, abduction block, headrest. It is highly important to secure patient because it safeguards him/her from falling out. If abduction contracture occurs, it is recommended to use abduction block in order to prevent ulcers between knees.
2. It is forbidden to use stroller for different purposes that it is intended to.
 3. It is forbidden to use stroller when it is damaged or any part is missing.
 4. It is forbidden to leave the patient unattended.
 5. Stroller should be used only when there is someone who is trained and knowledgeable.
 6. It is forbidden to leave stroller and the user on unstable ground unattended, even if the brake is engaged.
 7. It is forbidden to either go up or down the stairs; it doesn't matter if the user is placed in the device or not.





8. It is important to engage brake during getting into and getting out the stroller.
9. Make sure that the user does not stand on the footrest when getting in and out of the stroller.
10. Make sure that all limbs are not close to spokes or other adjustable elements.
11. The stroller should be stored and used away from sources of fire or high temperature. Moreover, stroller shouldn't be stored in rooms with high humidity.
12. Secure stroller during transportation in order to avoid any possible damages. Manufacturer doesn't take responsibility for damages which appeared due to improper preparation of the device during transportation.
13. Before using the stroller, remove all paper and plastic packaging. The packaging is not a toy! Keep away from children and animals.



It is forbidden to place the user in the device while being in car, bus or train.

The product contains small parts which children may swallow.

Do not use a product that has any signs of damage, tearing or wear. Only an undamaged product and proper use guarantee its right and safe working. If the product is damaged, discontinue use and contact the manufacturer.

Under no circumstances should an unqualified person repair or alter the product and its construction, as the health and safety of the user depend on it as well as it will lead to the loss of warranty.

Remember that all the loads hung on the push handle may cause the stroller to lose its balance.

The weight of user should not be bigger than maximal weight described on page 9 where you will find information regarding this issue.

Serious incidents involving this device should be reported both to manufacturer and to the competent authority in your country.





9. TERMS OF WARRANTY

- 1) **Akces-MED Ltd.** sells products free from manufacturing defects or faults.
- 2) Manufacturer gives the purchaser **24-month** guarantee from the date of purchase. All defects revealed during warranty period will be repaired free of charge.
- 3) All defects which appear after warranty period can be repaired only when the user will pay for it. This rule doesn't apply to refunded products - user can ask National Health Refund for refund.
- 3a) The stroller is covered by a 36-month warranty on the territory of Canada – the details of the rights resulting from the extended warranty are regulated by the warranty card.
- 4) Manufacturer claims that the product can be used for **48-months**. User can use this product after this period only when he/she will send the device to manufacturer so as to service it. Next, manufacturer determines again how long the product can be used and, moreover, the date of the next service will be determined.
- 5) If product is damaged, please contact our dealer or **Akces-MED Ltd.**
(tel. +48 17 864 04 77, e-mail: export@akces-med.com).

10. MAINTENANCE AND CLEANING

The stroller **TATALU™**, which is a medical device, should always be kept clean and should be used in accordance with the manufacturer's instructions. The stroller should be kept in a clean, dry and ventilated room. Maintenance and cleaning must be done at least once a month. Periodic cleaning and maintenance of the chair is essential to keep its proper functioning and safety.

Recommendations for cleaning the upholstery and other elements of the stroller:

1. Wash with warm water using a soft cloth and mild cleaners.
2. Make sure the cloth is not too wet, as this may enlarge the stain and damage the upholstery or foam inside the upholstery.
3. Remember to remove excess moisture from the upholstery before the next cleaning.
4. After cleaning the stains, the upholstery should be rinsed with lukewarm water and dried with a clean cloth.
5. You must not soak the upholstery.
6. The upholstery must not get wet during cleaning.
7. Do not use hard brushes for cleaning the upholstery.
8. Do not clean the upholstery with strong chemicals, e.g. solvents or chlorine-based caustics.





9. Do not clean, replace any elements or maintain the chair during use.
10. Metal and plastic elements should be cleaned by hand with suitable cleaners, after having read the recommendations of the manufacturer of the given cleaner.
11. It is possible to disinfect metal, plastic, wooden and upholstery elements of the product with a generally available disinfectant. In order to disinfect the selected element, spray its surface with a disinfectant generally available on the market and wipe with a dry, clean cloth. After disinfection, wait until the element is completely dry, especially in the case of upholstery. Before using the disinfectant, first check in an invisible place whether the product will not discolor or damage the disinfected surface.



Remember that upholstery has to be always dry before using!

The manufacturer of the stroller, **Akces-MED Ltd.**, is not responsible for any damages caused by using unsuitable cleaners.

Before using it is necessary to:

- clean wheels,
- check if base as well as wheels are correctly mounted, check if brake stops turning wheel,
- check if wheelchair construction is free of breaks or any kinds of deformities,
- check if all movable parts are working and all screws, nuts, knobs etc. are correctly tightened,
- check if belts, vests, trunk belts, feet stabilizing belts etc are not frayed; check if buckles and clamps work correctly.



Manufacturer takes responsibility only for hidden faults or damages which appeared despite the fact that user complied with the guidelines.

In case of any damages, it is essential to contact the person who sold you the product.





TECHNICAL INSPECTION

Technical inspections should be performed regularly. Before servicing, the product should be thoroughly cleaned to reveal any hidden damage.

| Frequency | What should be checked? What should be done? | | Remarks |
|------------------|--|---|--|
| Before each use. | General condition of the product. | Check: <ul style="list-style-type: none"> the entire product for any visible damage, cracks or mechanical defects,, if the product complete, if all bolts, screws, knobs, nuts and other elements of the construction are complete and tightened.. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Wheels and brakes. | Check: <ul style="list-style-type: none"> if the wheels are properly mounted and if they are not damaged, if the brake is blocking the spinning wheel, if the wheels are not dirty. If necessary, clean the wheels. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Fastening of the construction elements. | Check all the mechanisms of the product and its moving parts. Check in particular: <ul style="list-style-type: none"> if the seat is properly mounted (if applicable), if the wheels are properly mounted (if applicable), if the elements securing the user are properly mounted and if they work well (if applicable), if all the moving parts are complete, work properly, do not jam, if the adjustment mechanisms of the following elements work properly (if applicable): headrest, backrest, seat, abduction block, armrests, footrests. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Upholstery elements. | Check: <ul style="list-style-type: none"> if there is any damage, tears, rips, loose seams or deformations on the upholstery elements, if the stabilizing straps, vests or safety belts are not frayed or twisted and if the buckles work properly, if the stabilizing straps, vests or safety belts work correctly and do not unfasten under pressure or when pulled. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| Once a week. | Product cleaning. | If the frame and upholstery get dirty (dusty, muddy, etc.) during everyday use, wipe them with a damp cloth without detergents. Remove dirt (e.g. hair, food remains, etc.) from the moving parts of the product. | The product should be cleaned when necessary, but at least once a week. Do not use any chlorine based cleaners or methylated spirit. Before using the product, check if the upholstery is dry. |





| Frequency | What should be checked? What should be done? | | Remarks |
|--|---|---|--|
| Once a month. | Wheel axles, quick connectors (if applicable). | Check: <ul style="list-style-type: none"> if the wheels are rolling and if they are not loose. To maintain the wheels, apply a small amount of a preservative-lubricant between the axle and the wheel hub. Check if you can easily remove and attach the wheels. | Clean the wheels when dirty. In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Product construction connectors (welding, threading, soldering) | Check: <ul style="list-style-type: none"> the condition of the product frame in the places of joining (welds, threads, soldering). | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Tires/innertubes (if applicable). | Check: <ul style="list-style-type: none"> tire wear condition. Check for visible damage. Check tire pressure (if applicable). | If the tire pressure is too low, inflate them to the correct level. In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| Once every 6 months. | Construction moving parts. | Check: <ul style="list-style-type: none"> if the moving parts of the construction have not been damaged. Maintain them using a proper cleaning agent. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| | Product labeling. | Check: <ul style="list-style-type: none"> if the marking and labels are readable. | In case of any damage or defects, discontinue the use of the product and contact your Supplier/Distributor or the Manufacturer. |
| After the life time of the product finishes. | Detailed product check. | Check the product. | The manufacturer specifies the life time of the product, which is given in the user manual. After the given period, the user may use the product only if it is sent back to the manufacturer to be checked first. After that, the manufacturer will re-define the life time of the product and the date of the next check. |

The list of all spare parts is available for download after selecting a product on www.akces-med.com.





11. LABELLING

11.1 SYMBOLS

CE Declaration that the product meets the essential requirements

REF Catalog number

SN Serial number

Please read the manual instructions before use

You need to take special precautions

Manufacturer

Production date

The product does not contain latex

Store in a dry place

Prohibition of use without fastened safety belts

Protect from sunlight

Storage temperature limitation

Relative humidity storage

Dry vertically

Do not chlorinate

Do not iron

Do not dry in a drum dryer

Do not bleach

Do not chlorinate

Careful! Delicate content

Prohibition to leave the user without care during use of the product

It is forbidden to either go up or down the stairs; it doesn't matter if the user is placed in the device or not


You need to read the additional information

You need to press the button shown in the photo

You need to use the tools indicated in the assembly description

MD Medical device


11.2 SAMPLE DATA PLATE



MANUFACTURER



Jasionka 955B
36-002 Jasionka (POLAND)
tel. +48 17 864 04 77
e-mail: export@akces-med.com
BDO: 000031619



Authorised Dealer



akces-med.com

Date of issue: 30.03.2021
The last update: 22.09.2022
© Sławomir Wroński 2022