



DECLARATION OF EC CONFORMITY

Manufacturer Akces-MED Ltd.

Declare

that special needs stroller

HARMONY

BASIC UDI-DI: 59038165HRY34

of intended use: for people with musculoskeletal diseases/disabilities that prevent or limit their ability to stand upright and/or move independently.

Marked with the CE sign is Class I medical device, of rule 1, consistent with the requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.

The above product to which this declaration applies complies with Regulation (EU) 2017/745 of the European Parliament and of the Council.

Conformity assessment was carried out in accordance with the requirements of the Regulation (EU) 2017/745 of the European Parliament and of the Council.

In order to demonstrate the safety and effectiveness of the medical device, the following standards were used to assess conformity:

PN-EN ISO 20417:2021-10

Medical devices -- Information to be supplied by the manufacturer.

PN-EN ISO 21856:2023-01

Assistive products -- General requirements and test methods.

PN-EN ISO 15223-1:2022-01

Medical devices -- Symbols to be used with information to be supplied by the manufacturer

-- Part 1: General requirements.


PN-EN ISO 13485:2016-04

Medical devices -- Quality management systems -- Requirements for regulatory purposes.

PN-EN ISO 14971:2020-05

Medical devices -- Application of risk management to medical devices.

President of the board



Sławomir Wroński